

E06: Calibration Certification

Effective Date: 11/14/2019

The calibration supplier listed on the purchase order shall perform all calibrations and is not authorized to subcontract any calibration unless approved by Aerojet Rocketdyne. A calibration report shall be furnished to Aerojet for each calibration performed and at a minimum include the following information:

- Identification of the unit being calibrated
- Identification of the calibration source traceable to NIST
- Date of calibration
- Conditions "As Found" and "As Left"
- Notification of out-of-tolerance conditions
- Certification that calibrations were performed in accordance with either ANSI/NCSL Z540, ISO 10012 or ISO/IEC 17025 (latest revisions).
- Identification of measurement uncertainties